



### Declaration of Conformity

For the following equipment :

Product Name: Medical Type Switching Power Supply

Model Designation: MSP-450-x (x=3,3,5,7,5,12,15,24,36,48)

is herewith confirmed to comply with the requirements set out in the Council Directive 93/42/EEC concerning Medical devices, the following standards were applied :

**RoHS Directive (2011/65/EU), (EU)2015/863**

**MDR Directive (EU) 2017/745**

EN 60601-1:2006+A1:2013+A12:2014+A2:2021

CB certificate No : DK-150389-UL

EN 60601-1-2:2015+A1:2021

#### EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

	EN 55011:2016/A2:2021 (Group 1)	Class B
Harmonic current	EN IEC 61000-3-2:2019	
Voltage flicker	EN 61000-3-3:2013/A1:2019	

#### EMS (Electro-Magnetic Susceptibility)

ESD air	EN 61000-4-2:2009	Level 4	15KV
ESD contact	EN 61000-4-2:2009	Level 4	8KV
RF field susceptibility	EN IEC 61000-4-3:2020	Level 3	10V/m(80MHz-2.7GHz)
RF field susceptibility	EN IEC 61000-4-3:2020	Table 9	9~28V/m (385MHz~5.78GHz)
EFT bursts	EN 61000-4-4:2012	Level 3	2KV/100KHz
Surge susceptibility	EN 61000-4-5:2014/A1:2017	Level 4	2KV/Line-Line 4KV/Line-Earth
Conducted susceptibility	EN 61000-4-6:2014	Level 3	10V
Magnetic field immunity	EN 61000-4-8:2010	Level 4	30A/m
Voltage dip, interruption	EN IEC 61000-4-11:2020	100% dip 1 periods	30% dip 25 periods 100% interruptions 250 periods

#### Note:

A component power supply with load will be installed into final equipment which consists of an electronically shielded metal enclosure. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure.

For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on <http://www.meanwell.com>)".

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC4xxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

(Signature)

Alex Tsai/Director, Product Strategy Center :

(Name / Position)

(Signature)

Taiwan

(Place)

Feb. 16th, 2024

(Date)